Memorandum of Information TNO

Guidelines for the European Public procurement tender for the delivery of Scanning Electron Microscope Ref. 2022 FPL/INK 103 date February 23, 2023

Date :	**March 1, 2023**
Version:	**1**

Questions about the European Public procurement tender procedure and/or the European Public procurement tender documents can be asked until no later than the closing date of the 'Deadline for asking questions' as mentioned in the tendering schedule shown in sub-section 2.3.

Section nr. / subject	Questions Tenderer	Answer TNO
8.10 Sample preparation system	*Do sample 2 and 3 for the sample test require air isolation after milling? If that is required this will make scheduling the demo work a bit lex flexible. A quick reply ont his would be greatly appreciated.*	The samples provided can be imaged under normal conditions and do not need to be implemented to the SEM under airtight conditions.
8.10 Sample preparation system	*Does sample 1 require stage tilt or is using a cross section holder ok?*	Imaging of the CNT sample should not require stage tilt. Both surface imaging and cross section imaging don't need this.
9 Appendices	*Can the NDA be signed by someone with the foreign nationality?*	The NDA must be signed by a person who is legally authorized to sign on behalf of the tenderer. The signatory's authorization to sign must be established by means of a certified extract from the Chamber of Commerce Trade Register.

Section nr. / subject	Questions Tenderer	Answer TNO

Section nr. / subject	Questions Tenderer	Answer TNO

Section nr. / subject	Questions Tenderer	Answer TNO

Section nr. / subject	Questions Tenderer	Answer TNO