

**Aanbesteding:** FLI - BLI CT System  
**Aanbestedende Dienst:** Leids Universitair Medisch Centrum (LUMC)  
**Referentie:** L-EU-23-08

**Toelichting:**

**Ref.nr.**                      **Onderwerp:**  
1                                      Pos. 2.7.

**Vraag:**

You mention that "X-ray/Radiographic Imaging is available as optional and field upgradable image modality at no additional cost." The upgrade to X-Ray is a purchasable option. Can you explain this what do you mean with no additional cost?

**Antwoord:**

X-ray/Radiographic Imaging is a 'should/wish' and therefore optional. However, if you can deliver a device with X-ray/Radiographic Imaging and your answer to Pos. 2.7 is 'Yes', you must include these costs in the Price List. It is not permitted to charge additional costs for the X-ray/Radiographic Imaging apart from the costs stated in the Price List. Please note that the delivery of the X-ray/Radiographic Imaging modality must be within 6 months after ordering. However, it is not mandatory to deliver the X-Ray modality at the same time as the optical imaging modality, as long as the separate deliveries of both modalities take place within the 6-month period.

**Percelen:**                      P1 FLI-BLI CT System  
**Beantwoord op:**              21 sep. 2023  
**Label:**                              Inhoud

**Ref.nr.**                      **Onderwerp:**  
2                                      Schedule of Requirements Document, chapter 2.3.10

**Vraag:**

You mentioned In 2.3.10: The minimal optical FOV (field of view) is not less than 7.5 x 7.5 cm to obtain specimen imaging at highest possible resolutions.

It should be "less than 7.5 x 7.5cm" to reach a higher resolution.

**Antwoord:**

Correct, we have made a typographical error. The adjusted Requirement will be: The minimal optical FOV (field of view) is less than 7.5 x 7.5 cm to

obtain specimen imaging at highest possible resolutions.

**Percelen:** P1 FLI-BLI CT System

**Beantwoord op:** 21 sep. 2023

**Label:** Inhoud

**Ref.nr.**

3

**Onderwerp:**

Must have

**Vraag:**

In case we do not meet a 'must have' item, will this exclude us from the tender? Or we lose points only?

**Antwoord:**

The Tenderer its offer must unconditionally meet the requirements ('must have') set by the LUMC for the performance of the contract. This means that all requirements must be answered with "yes". When the LUMC concludes that the Tender does not comply with the requirements, the offer will be rejected and the Tenderer will be excluded from any further participation (knock-out).

**Percelen:** P1 FLI-BLI CT System

**Beantwoord op:** 21 sep. 2023

**Label:** Inhoud